



CERTIFICATE OF COMPLIANCE

Certificate Holder: LINAK Australia Pty Ltd

Manufacturer: LINAK A/S, Denmark

Certified Product: Actuator Control Box (for medical linear actuators)

Model: CO61

Brand: LINAK

Ratings: 100-240VAC, 50/60Hz, max. 3.9A, Class II, IPX6
No Applied Part, Duty Cycle: 10% max. 2 min/18 min

Test Specification: AS/NZS 60601.1: 2015
(Excluding clause 48 Biocompatibility, clause 14 PEES)
AS/NZS 3200.1.2: 2005

Test Report: UL Demko A/S E346122-D1005-1/A0/CO-ULCB
Delta T600363

Compliance Marking: CBA170360

This Certificate of Conformity was issued under the Rules of Certification for Electrical Product Safety Certification Scheme (EPSCS) of Certification Body Australia and it is based on a voluntary evaluation of a sample of the above mentioned product. Certification Body Australia certifies that the tested sample complies with all applicable requirements of the standards and regulations listed above. This certificate does not imply assessment of the production of the product and does not permit the use of the Certification Body Australia Certification Mark.

Certificate No.: CBA170360.01

Date of Issue: 27 April 2017

Valid until: 26 April 2022



810 Pacific Highway, Gordon NSW 2072, Australia

SOP2-6 2.0

Authorised Signatory:

A handwritten signature in blue ink, appearing to read 'Günther Theisz'.

Günther Theisz



www.jas-anz.org/register